

Test Verification of Conformity

Verification Number: 250300239TPE-001-VOC1

On the basis of the tests undertaken, the sample<s> of the below product has been tested by 3rd party laboratory in accordance to the referenced specification<s>/standard<s> at the time the tests were carried out. This verification is part of the full test report<s> and should be read in conjunction with it <them>.

Applicant Name & Address:	Bio-Helix Co Ltd. 5F., No. 145, Sec. 3, Beixin Rd., Xindian Dist., New Taipei City, 231030, Taiwan
Product Description:	UVPAD Trio LED 365/302/300nm UV Transilluminator
Ratings & Principal Characteristics:	PSU: I/P: 100-240 V~, 50/60 Hz, 1.5-0.8 A; O/P: 24 Vdc, 2.5 A, Class II; Main Unit: 24 Vdc, 2.5 A, Class III
Models/Type References:	UP001CU, UP001CUN
Brand Names:	Bio-Helix
Specification<s>/Standards:	EN 61010-1:2010/A1:2019; EN IEC 61010-2-101:2022/A11:2022; EN 62471: 2008
Verification Issuing Office Name & Address:	Intertek Testing Services Taiwan Ltd. 5F, No. 423, Ruiguang Rd., Neihu District, Taipei City 114, Taiwan
Date of Tests:	28 Apr. 2025 to 11 Jul. 2025
Test Report Number(s):	250300239TPE-001 and 250300240TPE-001



Signature

Name: Dan Chen
Position: Manager
Date: 26 Sep. 2025

This Verification is for the exclusive use of Intertek's client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this Verification. Only the Client is authorized to permit copying or distribution of this Verification. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test/inspection results referenced in this Verification are relevant only to the sample tested/inspected. This Verification by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.